



**HAL**  
open science

## High Speed Atomic Force Microscope

Nicolas Mauran, Denis Lagrange, Xavier Dollat, Laurent Mazonq, Lucien Schwab, Jean-Paul Salvetat, Bernard Legrand

► **To cite this version:**

Nicolas Mauran, Denis Lagrange, Xavier Dollat, Laurent Mazonq, Lucien Schwab, et al.. High Speed Atomic Force Microscope. NI-Week 2017, May 2017, Austin, TX, United States. , 2017. hal-01529673

**HAL Id: hal-01529673**

**<https://laas.hal.science/hal-01529673v1>**

Submitted on 31 May 2017


**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

## Summary


**Atomic Force Microscope (AFM) is now a common tool** for material analysis in the academic and industrial areas because it enables non-destructive high-resolution images of nanometric objects. However, a main drawback is the **slow scan rate** that hinders many potential applications. Recently, breakthroughs have been achieved in AFM sensors based on MEMS technology, allowing to extend AFM operation in terms of measurement bandwidth and data acquisition. The present work focusses on **developing an electronic controller** for AFM featuring the **wide bandwidth and the fast data processing rate** required to enable the exploitation of the full potential of MEMS AFM sensors.

## Microscope




Making of AFM microscope :  
 (a) and (e) : 3D drawing and photo of the microscope assembled.  
 (b) and (c) : design and making of fast Z actuator, wich consists of 2 piezoelectrics ceramics in opposite move, enclosed and glued between preload plates made of stainless steel.

## Electronics



Homemade electronic board to drive the fast Z actuator, consists of a 20-bit DAC and two (100V, 500 kHz) amplifiers to drive the two opposite ceramic piezos. The 20-bit word is generated by NI-9401 digital output module via SPI bus at 40MHz.

## Piezo Z control Loop timing



FPGA : microscope Z feedback operations take about 4 μs to complete. It executes every 1 μs with pipelining method.

## Software performance

	Small image 1k pixels	Biggest image 3M pixels
Scan duration	0.02s (50 images/s)	60s
Binary file size	47 KB	140 MB

## Software architecture

### PC – LabVIEW 2016, QMH project

**UI@ 10 ms**  
 R/W network streams from RT  
 Unbundle datas  
 Convert data to physical units (m,V)  
 Averaging per pixel

**UI@ 100 ms**  
 Display scaled data  
 Display scan images

**UI management**  
 Save image files  
 Control USB motor for approach

### Real-Time Controller – LabVIEW RT, QMH model

**RT@ 5 ms non determ.** → R/W stream to UI  
**RT@ 1 ms deterministic** → Read data from FPGA

### FPGA – LabVIEW FPGA

**Z control loop @1μs pipelining**  
 Probe signal acquisition  
 PID calculation with gain schedule  
 Piezo Z command (20-bit word to external DAC)

### Scan control @10μs or more


Scan X, Y via Analog Outputs  
 Read X, Y position sensors (AIs) → Watchdog @40MHz

## NI CompactRIO configuration





The controller of the AFM microscope is based on a 8-slot CompactRIO NI-9035.


Compact-RiO NI-9035	AIs NI-9215	AIs NI-9223	AOs NI-9263	DIOs NI-9401
	100 kSa/s 16 bits	1 MSa/s 16 bits	100 kSa/s 16 bits	




## 2-screen wide User Interface



Topography images of diblock copolymers in FM-AFM mode  
 (a) Image 1 × 1 μm (256 × 256 pix.) acquired in 60 s. (b) and (c) Images forward and backward 250 × 250 nm (128 × 128 pix.) acquired in 3 s.  
 The topographic contrast is due to elasticity contrast of copolymers strips.




Images of graphite HOPG steps in AM-AFM mode  
 (a) Image 2.5 × 2.5 μm (512 × 512 pix.) acquired in 130 s. (b) Image 1 × 1 μm (200 × 200 pix.) acquired in 40s. (c) Image 1 × 1 μm (200 × 200 pix.) acquired in 20 s.



Global view of High Speed AFM setup at LAAS-CNRS, Toulouse (France)

## High frequency AFM probes (MEMS technology)



www.vmicro.fr

AFM probes used in the present work were developed at IEMN-CNRS (Lille, FRANCE) and are now available from Vmicro SAS. A silicon ring holding a nanotip vibrates according to the elliptical resonance mode shape at about 13 MHz. **Capacitive electromechanical transducers** are integrated for driving and sensing the nanotip vibration. Typical measurement resolutions are 1.5 fm/√Hz in displacement and 0.5 pN/√Hz in force.